ΗΙΟΚΙ

INSULATION TESTER IR4056-20, IR4057-20

Field Measuring Instruments 7

5-range INSULATION & CONTINUITY

LINE

 \square

500V/1000V

RELEASE

vears

 \wedge

DROP -







Comparator Function Improves Work Efficiency

COMP

- Identify the Insulation and Low Resistance Conditions with the PASS/FAIL Icon
- FAIL Alert with Red LCD and Audio Buzzer

Stable Digital Readings are Easy to Read



LIGHT

CAT III 600V

IEC/EN 61557

EARTH


PASS COMP

HIOKI


0Ω ADJ



CE



Efficient, Safe Measurement with Digital Insulation Resistance Testers



Comparator function provides PASS/FAIL decisions at a glance

- The comparator function compares measured values to pre-set reference values to generate a pass or fail judgment. (Can be used with insulation resistance measurement and low-resistance measurement.)
- The stable display is easy to read, increasing work efficiency.

Instant judgment


Since the IR4056-20 and IR4057-20 generate judgments as soon as the test lead makes contact, it is possible to make a rapid series of measurements in the manner of a continuity check.



*In some cases, the capacitance component may prevent a judgment from being made until charging completes.

Identify PASS/FAIL using light and sound

The IR4056-20 and IR4057-20 notify the operator of pass and fail judgments using a beeping sound, LCD light, and comparator indicator on the test lead with remote control switch (optional accessory), allowing determinations of compliance to be made without looking at the instrument.



Designed for safety and peace of mind Featuring improved convenience and ease of use

DROP PROOF

Testers are built tough to withstand a 1-meter drop onto a concrete floor



AC/DC voltage measurement (With AC/DC automatic detection function)

Use as a tester replacement thanks to DC voltage measurement functionality, which is useful in applications involving solar power and electric vehicles (EVs).



200 mA grounding line continuity check function

The IR4056-20 and IR4057-20 can perform EV and HEV continuity checks as well as resistance measurement of safety conductors in building electrical equipment as defined by IEC 60364.



Safety-oriented double-action

500 V/1000 V range only





Press the flashing "RELEASE" key.



TEST LEAD L9787



Connect either the test probe or alligator clip for the earth side

Easy-to-see LCD

An FSTN LCD ensures the instrument's display is easy to read from any angle.

Effective maximum display value

A ">" mark is displayed when the measured value is greater than the effective maximum display value for the function in use.

Backlight (White LED)

A backlight makes it possible to work in dark or poorly lit locations.




IR4056-20 Economic model



5 ranges 50/125/250/500/ 1000 V	Rated output voltage (DC)	5 ranges 50/125/250/500/ 1000 V	
1	Voltage measurement	1	
1	Resistance measurement	1	
Approx. 0.8 s	Comparator judgment result response time	Approx. 0.3 s	
1	200 mA continuity	1	
-	Bar graph	✓	
159W×177H×53D	Dimensions(mm)	159W×177H×53D	
600	Mass(g)	640	


IR4057-20 Bar graph for visual judgments High-speed model

500





Bar graph



Specifications Guaranteed accuracy period: 1 year, Accuracy guarantee for temperature and humidity: 23°C±5°C (73°F ±9°F) and 90% RH

Insulation resistance measurement

Rated output voltage (DC)	50V	125V	250V	500V	1000V
Effective maximum indicated value	100 MΩ	250 MΩ	500 MΩ	2000 MΩ	4000 MΩ
Effective medium value	2 MΩ	5 MΩ	10 MΩ	50 MΩ	100 MΩ
1st effective measuring range $[M\Omega]$	0.200 to 10.00	0.200 to 25.0	0.200 to 50.0	0.200 to 500	0.200 to 1000
Accuracy	±4 % rdg.				
2st effective measuring range $[M\Omega]$	10.1 to 100.0	25.1 to 250	50.1 to 500	501 to 2000	1010 to 4000
Accuracy	±8 % rdg.				
Other measuring range $[M\Omega]$	0 to 0.199				
Accuracy	±2 % rdg. ±6 dgt.				
Lower limit resistance value to maintain nominal output voltage	0.05 MΩ	0.125 MΩ	0.25 MΩ	0.5 MΩ	1 MΩ
Overload protection	600 VAC (10 s)			1200 VAC (10 s)	

Test leads with sleeves

When measuring in a CAT IV or CAT III environment, be sure to attach the sleeve to the test leads. When the CAT (measurement category) rating of the main unit is lower than that of test leads, the CAT of the main unit takes precedence.



Basic specifications

Accessories

Instruction manual

LR6 alkaline battery × 4

HIOKI E.E. CORPORATION

TEST LEAD L9787

Neck strap

HEADQUARTERS:

Indicator: Semi-transmissive FSTN LCD, Positive backlight
Live circuit indicator, Automatic electric discharge, Automatic DC/AC detection, Comparator, Built-in battery power indicator etc.
Bar graph, Displaying 1-min. values
LR6 alkaline battery × 4
Approx. 20 hours (Comparator off, backlight off, 500 V range, no load)
The power will go off automatically 10 minutes after the last operation
0 to 40°C (32 to 104°F) 90% RH or lower (non-condensating)
-10 to 50°C (14 to 122°F) 90% RH or lower (non-condensating)
600 V AC/DC, Measurement category III, Anticipated transient overvoltage: 6000 V
7060 V AC, 50/60 Hz, Measurement terminals - electrical enclosure, 1 min, current sensitivity 1 mA
IP40
EN61326 (EMC), EN61557-1/-2/-4*/-10
On concrete: 1 m
Approx. 159W×177H×53D mm (6.26"W×6.97"H×2.09"D) (excluding protrusions)
IR4056-20: Approx. 600g (21.2 oz) IR4057-20: Approx. 640g (22.6 oz) (including battery, excluding test lead)
Test lead L9787 × 1, Neck strap × 1, Instruction manual × 1, LR6 alkaline battery × 4

* Subclause 4.3 of Part 4 (interchanging of test leads) is not applicable when L9788-10 is used.

Voltage measurement

-	i i i i i i i i i i i i i i i i i i i						
DC V	Display range (Auto range)	4.2 V	42 V	420 V	600 V		
	Maximum indicated value	4.200 V	42.00 V	420.0 V	750 V		
	Resolution	0.001 V	0.01 V	0.1 V	1 V		
	Accuracy	±1.3 % rdg. ±4 dgt. *					
AC V	Display range (Auto range)	420 V (Minimum indicated value: 30.0 V)		600 V			
	Maximum indicated value	420.0 V		750 V			
	Resolution	0.1 V		1 V			
	Accuracy	±2.3 % rdg. ±8 dgt. *					
	Measurement principles	Average responding type					
	Frequency range	50/60 Hz					
	C/DC automatic etection range	AC detected at 30 V or greater (50/60 Hz) (Pulsating currents with an overlapping AC component of 30 V or greater are detected as AC)					
	Effect of temperature	Measurement accuracy per $1^{\circ}C \times 0.1$ (Applicable to the operating temperature range other than 18 to 28°C)					

* Ranges in excess of 600 V are outside the accuracy guarantee.

Resistance measurement

Display range (Auto range)	10 Ω	100 Ω	1000 Ω
Maximum indicated value	10.00 Ω	100.0 Ω	1000 Ω
Resolution	0.01 Ω	0.1 Ω	1Ω
(after zero adjustment)	0 to 0.19Ω : ± 3 dgt. 0.20 to 10.00Ω : $\pm 3\%$ rdg. ± 2 dgt.	±3%rdg	. ±2dgt.
Measuring current	200 mA or more (at 6 Ω or less) (Display value before zero adjustment)		
Overload protection	600 VAC (10s, using fuse)		



Note: Company names and Product names appearing in this catalog are trademarks or registered trademarks of various companies.

DISTRIBUTED BY

HIOKI (Shanghai) SALES & TRADING CO., LTD.: TEL +86-21-63910090 FAX +86-21-63910360 http://www.hioki.cn / E-mail: info@hioki.com.cn

HIOKI INDIA PRIVATE LIMITED: TEL +91-124-6590210 FAX +91-124-6460113 E-mail: hioki@hioki.in

 HEADQuart Lens:
 E-Thail: ΠΙΟΚΙΨΠΟΚΙ.

 81 Koizumi, Ueda, Nagano, 386-1192, Japan
 FL +81-268-28-0568

 http://www.hioki.com / E-mail: os-com@hioki.co.jp
 HIOKI SINGAPORE PTE. LTD.:

 TEL +65-6634-7677
 FAX +65-6634-7477

 E-mail: info@hioki.com.sg
 E-mail: info@hioki.com.sg

 HIOKI USA CORPORATION:
 HIOKI KOREA CO., LTD.:

 TEL +1-609-409-9109
 FAX +1-609-409-9108
 TEL +82-42-936-1281

 http://www.hiokiusa.com / E-mail: hioki@hiokiusa.com
 TEL +82-42-936-1281
 FAX +82-42-936-1284

All information correct as of Oct. 1, 2013. All specifications are subject to change without notice.